



PANEL CUTOUT

### Filter Specifications

Max. Leakage Current (L-G) @ 120VAC/60Hz	Max. Leakage Current (L-G) @ 250VAC/50Hz	Min. common mode insertion loss (L-G) (dB in 50 ohm system) Freq. - MHz								Min. differential mode insertion loss (L-L) (dB in 50 ohm system) Freq. - MHz								
		.03	.10	.15	.50	1.0	3.0	5.0	10	30	.10	.15	.50	1.0	3.0	5.0	10	30
2.0 micro-amps	5.0 micro-amps	-	5	8	10	10	12	12	12	12	3	7	13	13	24	30	30	35

Note 1: Recommended diameter for thru-hole (screw/nut combination) hardware mounting method. Other methods will require holes with different dimensions and parameters.

THIS DOCUMENT AND ITS CONTENTS ARE THE PROPRIETARY CONFIDENTIAL PROPERTY OF INTERPOWER CORPORATION. NO USE OR COPIES MAY BE MADE WITHOUT WRITTEN PERMISSION FROM INTERPOWER CORPORATION.

#### DIMENSIONS IN MILLIMETERS

TITLE: Two Function Hospital Grade Power Entry Module

REV	CHANGE DESCRIPTION	DATE	INITIALS
B	Added Note 1.	3/28/12	DDF
A	Updated description in title block to say "Hospital Grade"	2/23/12	RLT

DRAFT BY: RLT DATE: 9-29-05

PART NUMBER: 83550020

DRAFTING APPROVAL: RLT DATE: 9-29-05

MATERIAL:

DO NOT SCALE

INTERPOWER CORPORATION

SHEET 1 OF 1