



Note 1: Recommended diameter for thru-hole (screw/nut combination) hardware mounting method. Other methods will require holes with different dimensions and parameters.

### Filter Specifications

Max. Leakage Current (L-G) @ 120VAC/60Hz	Max. Leakage Current (L-G) @ 250VAC/50Hz	Min. common mode insertion loss (L-G) (dB in 50 ohm system) Freq. - MHz	Min. differential mode insertion loss (L-L) (dB in 50 ohm system) Freq. - MHz
2.0 micro-amps	5.0 micro-amps	.03 .10 .15 .50 1.0 3.0 5.0 10 30	.10 .15 .50 1.0 3.0 5.0 10 30
		- 2 4 9 10 12 12 12 12	3 7 13 13 24 30 30 35

THIS DOCUMENT AND ITS CONTENTS ARE THE PROPRIETARY CONFIDENTIAL PROPERTY OF INTERPOWER CORPORATION. NO USE OR COPIES MAY BE MADE WITHOUT WRITTEN PERMISSION FROM INTERPOWER CORPORATION.

#### DIMENSIONS IN MILLIMETERS

TITLE: Two Function Hospital Grade Accessory Power Module

PART NUMBER: 83550040

REV	CHANGE DESCRIPTION	DATE	INITIALS
B	Added dimension information to cutoutscrew holes and Note 1.	4/09/12	DDF
A	Updated description in title block to say "Hospital Grade"	2/23/12	RLT

DRAFT BY: RLT DATE: 9-29-05

DRAFTING APPROVAL: RLT DATE: 9-29-05

MATERIAL:

DO NOT SCALE

SHEET 1 OF 1

INTERPOWER CORPORATION