

# DEMKO CERTIFICATE

**Certificate No.** D-07034-M1  
**Page** 1/3  
**Date of Issue** 2019-08-07

**Certificate Holder** Interpower Corp.  
100 Interpower Ave, PO Box 115, Oskaloosa, IA 52577, USA

**Manufacturer** Interpower Corporation  
100 Interpower Ave,  
PO Box 115, Oskaloosa, IA 52577, USA

**Production site** Interpower Corporation  
100 Interpower Ave, PO Box 115, Oskaloosa, IA 52577,  
USA

**Certified Product Model** Non-rewirable two-pole hospital plug with earth pin  
042

**Trademark**



**Rated Voltage / Frequency** 250 Vac

**Rated Current / Power** 16 A

**Insulation Class** -

**Degree of protection (IP)** 20

**Tested acc. to** DS 60884-2-D1:2017

**Test Report No.** 4788905414C issued on 2019-07-31

**Additional** The report was revised to include technical modifications

**Expire date** 2029-01-13

---

**Certification Manager**  
Jan-Erik Storgaard

This is to certify that representative sample(s) of the Product described herein ("Certified Product") have been investigated and found in compliance with the Standard(s) indicated on this Certificate, in accordance with the D Mark Requirements. The Designated Certificate holder is entitled to use the d or for cables <DEMKO> for the Certified Product manufactured at the production site(s) identified above, in accordance with the D Mark Service Agreement, including without limitation the D Mark Testing and Certification Services Service Terms. Only those Products bearing the D Mark should be considered as being covered by UL's D Mark Service. This Certificate shall remain valid through the expiration date, unless terminated earlier in accordance with the Service Agreement including without limitation if the Standard identified on this Certificate is amended or withdrawn prior the expiration date.

**Certification Body**

UL International Demko A/S, Borupvang 5A, DK-2750  
Ballerup, Denmark, Tel. +45 44 85 65 65,  
info.dk@ul.com  
www.ul-europe.com



# Appendix DEMKO CERTIFICATE

**Certificate No.** D-07034-M1  
**Page** 2/3  
**Date of Issue** 2019-08-07

Additional Information:  
For Class I equipment  
Standard Sheet: DK 2-8a

This certificate replaces previous certificate no. D-07034 issued on 2019-01-14, due to:  
-Changed the cable supplier, see test report for details  
-Delete the type of cable: H03VV-F 3G0.75 mm<sup>2</sup>

Model 042 with the following cable types:

H05VV-F 3G0,75 mm<sup>2</sup>, H05VV-F 3G1.0 mm<sup>2</sup> and H05VV-F 3G1.5 mm<sup>2</sup>  
made by Interpower Corp. - <HAR> approved.

---

**Certification Body**

UL International Demko A/S, Borupvang 5A, DK-2750  
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com  
[www.ul-europe.com](http://www.ul-europe.com)



# Appendix DEMKO CERTIFICATE

Certificate No. D-07034-M1  
Page 3/3  
Date of Issue 2019-08-07



---

**Certification Body**

UL International Demko A/S, Borupvang 5A, DK-2750  
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com  
[www.ul-europe.com](http://www.ul-europe.com)

