



Note 1: Recommended diameter for thru-hole (screw/nut combination) hardware mounting method. Other methods will require holes with different dimensions and parameters.

### Filter Specifications

Max. Leakage Current (L-G) @ 120VAC/60Hz	Max. Leakage Current (L-G) @ 250VAC/50Hz	Min. common mode insertion loss (L-G) (dB in 50 ohm system) Freq. - MHz								Min. differential mode insertion loss (L-L) (dB in 50 ohm system) Freq. - MHz								
		.03	.10	.15	.50	1.0	3.0	5.0	10	30	.10	.15	.50	1.0	3.0	5.0	10	30
2.0 micro-amps	5.0 micro-amps	-	5	8	10	10	12	12	12	12	3	7	13	13	24	30	30	35

THIS DOCUMENT AND ITS CONTENTS ARE THE PROPRIETARY CONFIDENTIAL PROPERTY OF INTERPOWER CORPORATION. NO USE OR COPIES MAY BE MADE WITHOUT WRITTEN PERMISSION FROM INTERPOWER CORPORATION.

#### DIMENSIONS IN MILLIMETERS

TITLE: Two Function Hospital Grade Accessory Power Module

PART NUMBER: 83550060

INTERPOWER CORPORATION

REV	CHANGE DESCRIPTION	DATE	INITIALS
B	Added dimension information to cutout screw holes and Note 1.	4/09/12	DDF
A	Updated description in title block to say "Hospital Grade"	2/23/12	RLT

DRAFT BY: RLT DATE: 9-29-05

DRAFTING APPROVAL: RLT DATE: 9-29-05

MATERIAL:

DO NOT SCALE

SHEET 1 OF 1