

# Certificate



for European Product Safety

Reference No. 1018515

## AC Power Source

Type designation 855217\*\*, 855222\*\*

Certificate holder **Interpower Corporation**  
100 Interpower Ave.  
Oskaloosa, IA 52577  
UNITED STATES

The product complies with the standard(s) EN 61010-1:2001

Date of expiry 13 July 2015

EU Directive information *The product satisfies the provisions for CE marking according to the Low Voltage Directive 2006/95/EC.*

Additional information in Appendix.

Certification Body Intertek Semko AB, Product Certification

Place Kista - Stockholm

Signed

  
Bo Berglöv

Date 13 July 2010

Page 1 of 2



**Intertek  
Appendix**

Reference No. 1018515

*Technical data*

Type designation 855217\*\*, 855222\*\*  
Input 100-240V AC, 16A, 50-60Hz  
Output 16A max low range, 8A max. high range  
Class I  
Brand name:



*Product information*

The symbol "\*\*\*" denotes model variants with no safety critical differences.

*Manufacturing site(s)*

Interpower Corporation  
2900 South East 5th Street,  
Ames, IA 50010,  
UNITED STATES

This certificate, which includes production control, is considered to constitute sufficient support for an EC Declaration of Conformity and CE marking of the products according to the Low Voltage Directive 2006/95/EC.

If the product is to be marketed in different European countries, it must be provided with a certified plug appropriate for each country.

Safety related parts of the user instructions shall be written in a language acceptable to the country where the product is to be used.

*Certification Body* Intertek Semko AB, Product Certification *Place* Kista - Stockholm

*Signed*

  
Bo Berglöv

*Date* 13 July 2010

*This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.*